

1. This document was created to support maximum accessibility for all learners. If you would like to print a hard copy of this document, please follow the general instructions below to print multiple slides on a single page or in black and white.
2. If you are viewing this course as a recorded course after the live webinar, you can use the scroll bar at the bottom of the player window to pause and navigate the course.
3. This handout is for reference only. Non-essential images have been removed for your convenience. Any links included in the handout are current at the time of the live webinar, but are subject to change and may not be current at a later date.
4. Copyright: Images used in this course are used in compliance with copyright laws and where required, permission has been secured to use the images in this course. All use of these images outside of this course may be in violation of copyright laws and is strictly prohibited.

How to print Handouts

- On a PC
 - Open PDF
 - Click Print
 - Choose # of pages per sheet from dropdown menu
 - Choose Black and White from “Color” dropdown
- On a Mac
 - Open PDF in Preview
 - Click File
 - Click Print
 - Click dropdown menu on the right “preview”
 - Click layout
- Choose # of pages per sheet from dropdown menu
- Checkmark Black & White if wanted.
- If more details needed please visit our FAQ page: <https://www.audiologyonline.com/help>

No part of the materials available through the continued.com site may be copied, photocopied, reproduced, translated or reduced to any electronic medium or machine-readable form, in whole or in part, without prior written consent of continued.com, LLC. Any other reproduction in any form without such written permission is prohibited. All materials contained on this site are protected by United States copyright law and may not be reproduced, distributed, transmitted, displayed, published or broadcast without the prior written permission of continued.com, LLC. Users must not access or use for any commercial purposes any part of the site or any services or materials available through the site.

Technical issues with the Recording?

- Clear browser cache using [these instructions](#)
- Switch to another browser
- Use a hardwired Internet connection
- Restart your computer/device

Still having issues?

- Call 800-753-2160 (M-F, 8 AM-8 PM ET)
- Email customerservice@AudiologyOnline.com



GSI AMTAS: Implementation and Interpretation Case Studies





GSI Audiology Team

- Application Support
- Training
- R&D testing

Tony Lombardo



Laura Prigge



Karen Morris



Learning Outcomes

- 1) After this course, the participant will be able to use the quality indicators to interpret the quality of the automated audiogram.
- 2) After this course, the participant will be able to use the speech results to interpret the quality of the automated speech testing.
- 3) After this course, the participant will be able to determine most appropriate next steps for patients based on the results of the automated audiometry.



Are you “on the fence” when it comes to Automated Audiometry?



Brief History of Automated Audiometry

- Bekesy Audiometry
 - Considered Special Test
 - Not used routinely
- Military Annual Screening
- Hearing Conservation Programs
- Research Programs

Historically Screening: Confirm normal or refer for more testing. Air conduction only.
Little diagnostic relevance.



Current Status of Automation


- Technology advancements in medical professions
 - Blood pressure
 - Vision testing
 - Pure tone audiometry
- Provides continuous measurements
- Analyze and quantify behavioral responses
- Correlate responses for test accuracy




The Access Factor



Landscape of Audiology in the US




In the 2013-17 period, the proportion of older adults (≥ 65) self-reporting difficulty hearing ranged from 3.2% to 44.4%.



Data Source: US Census Bureau, ACS 2013-17, Table S11

Estimated percentage of county population aged ≥ 65 reporting hearing difficulty (data source: U.S. Census Bureau, American community survey Table S1810).

Audiologists per 100,000 Population (2015)



Increasing need for hearing tests in the US

- Population 2020-2060
 - 333 Million → 417 Million
- Prevalence of hearing impairment
 - 17% → 18%
 - 56 Million → 75 Million
- Number of hearing tests per patient per year
 - Assuming bi-annual evaluation
 - Minus 25% total tests for normal hearing
- Total Number of hearing tests needed per year
- 35 Million → 47 Million





What does this mean?


- Number of FT Audiologists
 - 12, 019 → 12, 019
- Number of FT HAD/HIS
 - 6,000 → 6,000
- Number of Hearing Tests Per Year Per Tester
 - 3 per workday
 - 920 per year
 - 17 Million → 17 Million



Unable to meet demand.




- Disabling Hearing Loss




World Health Organization. **Multi-country assessment of national capacity to provide hearing care.** ISBN 978 92 4 150657 1 (NLM classification: WV 270) © World Health Organization 2013

- Audiologists per million



ENT per million



How do we address this increasing need?

- Increase the workforce
- TeleMedicine/TeleAudiology
- More efficient use of the workforce
 - Use of technicians
 - Audiology Assistants



How do we address this increasing need?

- Automation
 - Pure tone air/bone
 - Speech
 - Basic Diagnostic tests



Brief History of Pure Tone Audiometry

- 1924: First pure tone methods published in the *Laryngoscope*
- 1944: Original Hughson-Westlake Method article published
- 1959: Carhart and Jerger method published
- 1970: Adaptive psychophysical procedures (used in research)
- 2004: ANSI Standards guidance on pure tone audiometry

leaves many procedural variables to the discretion of the tester

Manual audiometry has persisted without much change for 90 years




How consistent are pure tone tests from clinician to clinician?

How much variability is there in pure tone testing?



What if a patient brings in a recent audiogram from another clinic?





It is very common for the audio to be repeated because of lack of confidence in the original results.



The Human Factor – Tester Bias

- Experienced audiologists learn certain tricks that ensure accuracy.
- The varying skill levels of testers and their testing methods result in widely varying accuracy.
- Testers are sensitive to different cues that suggest that the patient may not be responding consistently and accurately.
- Experienced audiologists can basically predict the audio based on the first few responses and patient history.
- Experienced audiologists have the confidence to cut corners without affecting the integrity of the results.



Is there a standardized method for pure tone audiometry?

not as consistent as we thought



The Missing Link: Standardization

- The basic diagnostic eval does not require skillset of an expert audiologist
- “tell the story” of patient behavior during evaluation
- AMTAS provides patented Quality indicators to assist in interpretation
- Patented and validated



Standardized Automated Audiometry



Qualind™: A Method for Assessing the Accuracy of Automated Tests. Margolis, et al, 2007, JAAA, 18: 78-89.

- Patented method for determining the accuracy of a test results
- Data was collected at three sites from a wide range of settings, patient demographics, and hearing loss characteristics.
- Large subject sample (n = 120), a strong relationship was found between predicted and measured accuracy.
- Method may be useful for automated test procedures when skilled professionals are not available to provide quality assurance.



9 Quality Indicators (Qualind)

QUALITY ASSESSMENT (Air & Bone)			
①	Predicted Accuracy	GOOD	
	Quality Indicators	Value	File
②	Predicted Avg. Abs. Diff. (dB)	5.287	0
③	Masker Alert (%)	0.000	0
④	Time per Trial (s)	2.698	0
⑤	False Alarm (%)	3.846	35
⑥	Avg. Test Retest Diff. (dB)	5.000	46
⑦	QC Fail (%)	0.000	0
⑧	ABG > 35	0.000	0
⑨	ABG < -10	0.000	0

LEGEND		Right	Left
Masked Air	△		□
Masked Bone	▽		□

- **Predicted Accuracy** (Overall Quality -Good, Fair, Poor)
- **Predicted Average Absolute Difference** –Difference between automated and manual thresholds
- **Masker Alerts**-thresholds where masking may have been too high or low
- **Time per Trial** - average time from stimulus to patient response
- **False Alarm Rate** – number of times patient responded with no stimulus presented divided by number of trials when no stim present
- **Average Test-Retest Difference** – average difference between 1 KHz test and retest in right and left hear
- **Quality Check Fail Rate** – number of times patient did not respond to stimulus above threshold divided by the number of measured thresholds
- **Number of Air / Bone Gap > 35 dB** – number of air bone gaps that exceed 35 dB
- **Number of Air / Bone Gap < -10 dB** – number of air bone gaps that are less than 10db



9 Quality Indicators (Qualind)

QUALITY ASSESSMENT (Air & Bone)			
①	Predicted Accuracy	GOOD	
	Quality Indicators	Value	%ile
②	Predicted Avg. Abs. Diff. (dB)	5.287	0
③	Masker Alert (%)	0.000	0
④	Time per Trial (s)	2.698	0
⑤	False Alarm (%)	3.846	35
⑥	Avg. Test Retest Diff. (dB)	5.000	46
⑦	QC Fail (%)	0.000	0
⑧	ABG > 35	0.000	0
⑨	ABG < -10	0.000	0

LEGEND		Right	Left
Masked Air	△		□
Masked Bone	▽		□

- **Predicted Accuracy** (Overall Quality - Good, Fair, Poor)
 - “reliability” judgment found on audiogram forms assumes that a skilled audiologist can observe certain patient behaviors and characteristics that predict accuracy.



9 Quality Indicators (Qualind)

QUALITY ASSESSMENT (Air & Bone)		
①	Predicted Accuracy	GOOD
	Quality Indicators	Value File
②	Predicted Avg. Abs. Diff. (dB)	5.287 0
③	Masker Alert (%)	0.000 0
④	Time per Trial (s)	2.698 0
⑤	False Alarm (%)	3.846 35
⑥	Avg. Test Retest Diff. (dB)	5.000 46
⑦	QC Fail (%)	0.000 0
⑧	ABG > 35	0.000 0
⑨	ABG < -10	0.000 0

LEGEND		Right	Left
Masked Air	△		□
Masked Bone	▽		□

- **Predicted Average Absolute Difference**
–Difference between automated and manual thresholds
- Derived from study 2 data
- Helps establish confidence in the patient responses and accuracy of thresholds.



Study 1: Inter-tester Differences for Manual Audiometry

- 3 sites; 6 patients (included normal and sensorineural hearing loss) tested by 2 audiologists (with at least 10 years of experience)
 - Air-conduction thresholds absolute differences ranged from 2.4 to 6.3 dB. The overall mean absolute difference for all frequencies was 4.1 dB, with a 95% confidence interval of 2.3 to 6.0 dB.
 - For bone-conduction thresholds, mean absolute difference for frequencies tested, excluding 250 Hz, was 5.4 dB, with a 95% confidence interval of 2.9 to 7.9 dB.



Study 2: AMTAS Manual Threshold Comparison

- 30 subjects (5 normal hearing, 25 hearing impaired) were tested with AMTAS and manual audiometry
- Air-conduction absolute differences (AMTAS-Manual), ranged from 3.0 to 4.5 dB with mean absolute difference 3.6 dB.
- Bone-conduction thresholds (AMTAS-Manual), ranged from 7.4 to 8.1 dB with mean absolute difference of 7.7 dB. (Differences attributed to forehead vs mastoid placement and calibration)



9 Quality Indicators (Qualind)

QUALITY ASSESSMENT (Air & Bone)			
①	Predicted Accuracy	GOOD	
	Quality Indicators	Value	%ile
②	Predicted Avg. Abs. Diff. (dB)	5.287	0
③	Masker Alert (%)	0.000	0
④	Time per Trial (s)	2.698	0
⑤	False Alarm (%)	3.846	35
⑥	Avg. Test Retest Diff. (dB)	5.000	46
⑦	QC Fail (%)	0.000	0
⑧	ABG > 35	0.000	0
⑨	ABG < -10	0.000	0

LEGEND		Right	Left
Masked Air	△		□
Masked Bone	▽		□

- **Masker Alert**—thresholds where masking may have been too high or low
- Masking presented on every presentation
- Indicates over masking or under masking



9 Quality Indicators (Qualind)

QUALITY ASSESSMENT (Air & Bone)		
① Predicted Accuracy	GOOD	
Quality Indicators	Value	%ile
② Predicted Avg. Abs. Diff. (dB)	5.287	0
③ Masker Alert (%)	0.000	0
④ Time per Trial (s)	2.698	0
⑤ False Alarm (%)	3.846	35
⑥ Avg. Test Retest Diff. (dB)	5.000	46
⑦ QC Fail (%)	0.000	0
⑧ ABG > 35	0.000	0
⑨ ABG < -10	0.000	0
LEGEND		
	Right	Left
Masked Air	△	□
Masked Bone	▽	■

- Time Per Trial—average time from stimulus to patient response



9 Quality Indicators (Qualind)

QUALITY ASSESSMENT (Air & Bone)		
①	Predicted Accuracy	GOOD
	Quality Indicators	Value File
②	Predicted Avg. Abs. Diff. (dB)	5.287 0
③	Masker Alert (%)	0.000 0
④	Time per Trial (s)	2.698 0
⑤	False Alarm (%)	3.846 35
⑥	Avg. Test Retest Diff. (dB)	5.000 46
⑦	QC Fail (%)	0.000 0
⑧	ABG > 35	0.000 0
⑨	ABG < -10	0.000 0

LEGEND		Right	Left
Masked Air	△		□
Masked Bone	▽		□

- False Alarm Rate

- Catch trials are presented randomly throughout the test to allow a quantitative measurement of false-alarm rate.
- Patients can get into a “rhythm”
- Software delivers a message designed to keep them on task



9 Quality Indicators (Qualind)

QUALITY ASSESSMENT (Air & Bone)		
①	Predicted Accuracy	GOOD
	Quality Indicators	Value File
②	Predicted Avg. Abs. Diff. (dB)	5.287 0
③	Masker Alert (%)	0.000 0
④	Time per Trial (s)	2.698 0
⑤	False Alarm (%)	3.846 35
⑥	Avg. Test Retest Diff. (dB)	5.000 46
⑦	QC Fail (%)	0.000 0
⑧	ABG > 35	0.000 0
⑨	ABG < -10	0.000 0

LEGEND		Right	Left
Masked Air	△		□
Masked Bone	▽		□

- **Avg Test-Retest Difference**—average difference between 1 KHz test and retest in right and left hear.



9 Quality Indicators (Qualind)

QUALITY ASSESSMENT (Air & Bone)		
①	Predicted Accuracy	GOOD
	Quality Indicators	Value File
②	Predicted Avg. Abs. Diff. (dB)	5.287 0
③	Masker Alert (%)	0.000 0
④	Time per Trial (s)	2.698 0
⑤	False Alarm (%)	3.846 35
⑥	Avg. Test Retest Diff. (dB)	5.000 46
⑦	QC Fail (%)	0.000 0
⑧	ABG > 35	0.000 0
⑨	ABG < -10	0.000 0

LEGEND		Right	Left
Masked Air	△		□
Masked Bone	▽		□

• Quality Control Fail Rate

- Number of times patient did not respond to stimulus above threshold divided by the number of measured thresholds
- QC check is performed after each threshold determination by presenting a stimulus with a higher intensity than the threshold level.
- A “No” response to that stimulus is a “Quality Check Fail.”



9 Quality Indicators (Qualind)

QUALITY ASSESSMENT (Air & Bone)			
①	Predicted Accuracy	GOOD	
	Quality Indicators	Value	File
②	Predicted Avg. Abs. Diff. (dB)	5.287	0
③	Masker Alert (%)	0.000	0
④	Time per Trial (s)	2.698	0
⑤	False Alarm (%)	3.846	35
⑥	Avg. Test Retest Diff. (dB)	5.000	46
⑦	QC Fail (%)	0.000	0
⑧	ABG > 35	0.000	0
⑨	ABG < -10	0.000	0

LEGEND		Right	Left
Masked Air	△		□
Masked Bone	▽		□

- Air Bone Gaps–
- Number of Air / Bone Gap > 35 dB – number of air bone gaps that exceed 35 dB
- Number of Air / Bone Gap < -10 dB – number of air bone gaps that are less than 10db



9 Quality Indicators (Qualind)

QUALITY ASSESSMENT (Air & Bone)			
①	Predicted Accuracy	GOOD	
	Quality Indicators	Value	File
②	Predicted Avg. Abs. Diff. (dB)	5.287	0
③	Masker Alert (%)	0.000	0
④	Time per Trial (s)	2.698	0
⑤	False Alarm (%)	3.846	35
⑥	Avg. Test Retest Diff. (dB)	5.000	46
⑦	QC Fail (%)	0.000	0
⑧	ABG > 35	0.000	0
⑨	ABG < -10	0.000	0

LEGEND		Right	Left
Masked Air	△		□
Masked Bone	▽		□

- **Predicted Accuracy** (Overall Quality -Good, Fair, Poor)
- **Predicted Average Absolute Difference** –Difference between automated and manual thresholds
- **Masker Alerts**-thresholds where masking may have been too high or low
- **Time per Trial** - average time from stimulus to patient response
- **False Alarm Rate** – number of times patient responded with no stimulus presented divided by number of trials when no stim present
- **Average Test-Retest Difference** – average difference between 1 KHz test and retest in right and left hear
- **Quality Check Fail Rate** – number of times patient did not respond to stimulus above threshold divided by the number of measured thresholds
- **Number of Air / Bone Gap > 35 dB** – number of air bone gaps that exceed 35 dB
- **Number of Air / Bone Gap < -10 dB** – number of air bone gaps that are less than 10db




Audiogram Classification (AMCLASS)

- Audiogram is given three descriptive categories based on
 - Configuration: Normal, Flat, Sloping Hearing Loss, Rising, Trough, Peaked, and other
 - Severity: Normal, Mild, Moderate, Severe, Profound
 - Site of Lesion: Conductive, Sensory, Mixed
- Audiogram is analyzed for interaural asymmetry.



How does AMTAS automate speech testing?

- Forced Choice
- SRT
- WRS
- Task becomes TOO easy




Case Studies

- All submitted from actual audiologists using AMTAS
- Variety of settings
 - Large medical practice
 - ENT practice
 - Private practice with medical referrals



GSI AMTAS™ Hearing Report

AUDIOGRAM Frequency (Hz)



air	-10	-10	-10	0	20	20	RT
bone				50	75		RT
air	-10	-10	-5	15	20	15	LF
bone			40	50	75		LF

Effective Threshold Levels (dB HL)


QUALIND®		
Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	10
Masker Alert	0 %	0
Trial Time	3.0 s	60
False Alarm	3 %	25
Test-Retest	3 dB	23
Check Fail	0 %	0
ABG > 30	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	└	┐

QUALIND®		
Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	10
Masker Alert	0 %	0
Trial Time	3.0 s	60
False Alarm	3 %	25
Test-Retest	3 dB	23
Check Fail	0 %	0
ABG > 30	0	0
ABG < -10	0	0

GSI AMTAS™ Hearing Report

AUDIOGRAM Frequency (Hz)



Effective Masking Levels (dB EML)

QUALIND®		
Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	5
Masker Alert	0 %	0
Trial Time	3.9 s	65
False Alarm	0 %	0
Test-Retest	3 dB	23
Check Fail	0 %	0
ABG > 35	0	0
ABG < -10	0	0


LEGEND	Right	Left
Masked Air	△	□
Masked Bone	▽	◻

AMCLASS® Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	MILD	Flat Hearing Loss	SENSORINEURAL	Symmetric
Left	NORMAL-MILD	Sloping Hearing Loss	SENSORINEURAL	

Speech Recognition Report						
Ear	PTA dB HL	SRT dB HL	Word Recognition (NU6 closed set) Correct %	Level dB HL	Mask dB HL	List
Right	5	-5	88	27	7	NU6-4
Left	5	-5	88	27	7	NU6-2

QUALIND®		
Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	5
Masker Alert	0 %	0
Trial Time	3.9 s	65
False Alarm	0 %	0
Test-Retest	3 dB	23
Check Fail	0 %	0
ABG > 35	0	0
ABG < -10	0	0

GSI AMTAS™ Hearing Report




QUALIND®		
Quality	GOOD	
Indicators	Value	Rank
Accuracy	0 dB	10
Masker Alert	0 X	0
Trial Time	3.1 s	10
False Alarm	5 X	40
Test-Retest	3 dB	20
Check Fail	0 X	0
ABG > 35	0	0
ABG < -10	1	50

LEGEND	Right/Left
Masked Air	△ □
Masked Bone	▽ ▢


AMCLASS® Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	NORMAL		Cannot Determine	Symmetric
Left	NORMAL		Cannot Determine	

Speech Recognition Report						
Ear	PTA	SRT	Word Recognition (NIG closed set) List			
	dB HL	dB HL	Correct %	Level dB HL	Mask dB HL	NIG-1
Right	10	5	100	65	45	NIG-2
Left	12	5	94	65	45	NIG-1


Correction added to bone threshold at 4000 Hz.: 15 dB.




Speech	SDT		SRT		WRS / SRS 1				WRS / SRS 2				MCL		UCL	
	dB HL	(m)	dB HL	(m)	%	dB HL	(m)	S/N	%	dB HL	(m)	S/N	dB HL	dB HL		
Right			5		100.0	45										
Left			5		100.0	45										
Bin																
Note	1 NU-6 List 2A						2 NU-6 List 2A									
Aided																
Note	1						2									



Report Comments:



GSI AMTAS™ Hearing Report




QUALIND®		
Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	30
Masker Alert	0 X	0
Trial Time	4.2 s	75
False Alarm	4 X	35
Test-Retest	8 dB	70
Check Fail	0 X	0
ABG > 35	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	▽	◇


AMCLASS® Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	NORMAL		Cannot Determine	
Left	NORMAL-MILD	Sloping Hearing Loss	SENSORINEURAL	Symmetric

PTA	
Ear	dB HL
Right	7
Left	5


Correction added to bone threshold at 4000 Hz.: 15 dB.




Speech	SDT	SRT	WRS / SRS 1	WRS / SRS 2	MCL	UCL
	dB HL (m)	dB HL (m)	%	dB HL (m) S/N %	dB HL (m) S/N %	dB HL
Right		5	92.0	45		
Left		5	88.0	45		
Bin.						
Note	1 NU-6 List 3A			2 NU-6 List 2A		
Aided						
Note	1			2		



Report Comments:



GSI AMTAS™ Hearing Report



QUALIND®		
Quality	GOOD	
Indicators	Value	Rank
Accuracy	5 dB	0
Masker Alert	0 X	0
Trial Time	3.4 s	30
False Alarm	0 X	0
Test-Retest	3 dB	23
Check Fail	0 X	0
ABG > 35	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	▽	◇


AMCLASS® Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	NORMAL		Cannot determine	Symmetric
Left	NORMAL		Cannot determine	

Speech Description Based

QUALIND®		
Quality	GOOD	
Indicators	Value	Rank
Accuracy	5 dB	0
Masker Alert	0 X	0
Trial Time	3.4 s	30
False Alarm	0 X	0
Test-Retest	3 dB	23
Check Fail	0 X	0
ABG > 35	0	0
ABG < -10	0	0

GSI AMTAS™ Hearing Report

AUDIOGRAM Frequency (Hz)



air	-10	-5	-5	0	15	15	rt
bone		45	40	50	65		rt
air	-10	-5	-5	-5	0	20	lf
bone		45	35	40	70		lf

Effective Masking Levels (dB ENL)

QUALIND®

Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	20
Masker Alert	0 %	0
Trial Time	3.4 s	30
False Alarm	9 %	65
Test-Retest	3 dB	23
Check Fail	0 %	0
NO > 35	0	0
NO < -10	0	0

Legend	Right	Left
Masked Air	△	□
Masked Bone	▽	◇

QUALIND®


Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	20
Masker Alert	0 %	0
Trial Time	3.4 s	30
False Alarm	9 %	65
Test-Retest	3 dB	23
Check Fail	0 %	0
NO > 35	0	0
NO < -10	0	0

AMCLASS® Audiogram Classification

Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	NORMAL-MILD	Sloping Hearing Loss	SENSORINEURAL	Symmetric
Left	MILD	Flat Hearing Loss	SENSORINEURAL	

GSI AMTAS™ Hearing Report

AUDIOGRAM Frequency (Hz)



QUALIND®		
Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	5
Masker Alert	5 %	70
Trial Time	4 s	70
False Alarm	0 %	0
Test-Retest	0 dB	0
Check Fail	0 %	0
ABG > 35	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	∩	∪
Over Masking		

AMCLASS® Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	NORMAL-MODERATE	Sloping Hearing Loss	SENSORINEURAL	Asymmetric
Left	NORMAL-SEVERE	Sloping Hearing Loss	MIXED	


Speech Recognition Report										
Ear	PTA dB HL	SRT dB HL	Word Recognition (NU6 closed set)				Word Recognition (NU6 closed set)			
			Correct %	Level dB HL	Mask dB HL	List	Correct %	Level dB HL	Mask dB HL	List
Right	12	7	76	34	14	NU6-2	84	42	22	NU6-4
Left	15	15	64	37	17	NU6-1	76	45	25	NU6-3

QUALIND®

Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	5
Masker Alert	5 %	70
Trial Time	4 s	70
False Alarm	0 %	0
Test-Retest	0 dB	0
Check Fail	0 %	0
ABG > 35	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	∩	∪

GSI AMTAS™ Hearing Report



QUALIND®		
Quality	GOOD	
Indicators	Value	Rank
Accuracy	7 dB	35
Masker Alert	0 %	0
Trial Time	3.4 s	35
False Alarm	14 %	80
Test-Retest	3 dB	23
Check Fail	0 %	0
ABG > 35	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	▽	◇


AMCLASS® Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	NORMAL-MODERATE	Sloping Hearing Loss	SENSORINEURAL	Asymmetric
Left	NORMAL-MODERATE	Sloping Hearing Loss	SENSORINEURAL	

Speech Recognition Report										
Ear	PTA	SRT	Word Recognition (NU6 closed set)				Word Recognition (NU6 closed set)			
	dB HL	dB HL	Correct %	Level dB HL	Mask dB HL	List	Correct %	Level dB HL	Mask dB HL	List
Right	5	-5	80	27	7	NU6-4	88	35	15	NU6-2
Left	5	-5	84	27	7	NU6-1	96	35	15	NU6-3

QUALIND®		
Quality	GOOD	
Indicators	Value	Rank
Accuracy	7 dB	35
Masker Alert	0 %	0
Trial Time	3.4 s	35
False Alarm	14 %	80
Test-Retest	3 dB	23
Check Fail	0 %	0
ABG > 35	0	0
ABG < -10	0	0

Speech Recognition Report										
Ear	PTA	SRT	Word Recognition (NU6 closed set)				Word Recognition (NU6 closed set)			
	dB HL	dB HL	Correct %	Level dB HL	Mask dB HL	List	Correct %	Level dB HL	Mask dB HL	List
Right	5	-5	80	27	7	NU6-4	88	35	15	NU6-2
Left	5	-5	84	27	7	NU6-1	96	35	15	NU6-3

GSI AMTAS™ Hearing Report



QUALIND®		
Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	30
Trial Time	4.4 s	85
False Alarm	4 %	40
Test-Retest	3 dB	23
Check Fail	0 %	0

LEGEND	Right	Left
Masked Air	△	□


QUALIND®		
Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	30
Trial Time	4.4 s	85
False Alarm	4 %	40
Test-Retest	3 dB	23
Check Fail	0 %	0

AMCLASS® Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	NORMAL		Cannot determine	Symmetric
Left	NORMAL		Cannot determine	

Speech Recognition Report						
Ear	PTA dB HL	SRT dB HL	Word Recognition (NU6 closed set)			
			Correct %	Level dB HL	Mask dB HL	List
Right	0	-5	100	22	2	NU6-3
Left	5	-5	92	27	7	NU6-4

GSI AMTAS™ Hearing Report

AUDIOGRAM
Frequency (Hz)



air	-10	-10	-10	-10	-5	15	Rt
bone					15		Rt
air	60	60	60	55	60	50	Lf
bone		80	80	80	90		Lf

Effective Masking Levels (dB EML)

AMCLASS® Audiogram Classification

Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	MILD	Flat Hearing Loss	SENSORINEURAL	Asymmetric
Left	SEVERE	Flat Hearing Loss	SENSORINEURAL or MIXED	

QUALIND®

Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	5
Masker Alert	16 %	85
Trial Time	3.2 s	15
False Alarm	0 %	0
Test-Retest	0 dB	0
Check Fail	0 %	0
ABG > 35	0	0
ABG < -10	0	0


LEGEND	Right	Left
Masked Air	△	□
Masked Bone	∟	∟
NR	↖	↗
Under Masking		

QUALIND®

Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	5
Masker Alert	16 %	85
Trial Time	3.2 s	15
False Alarm	0 %	0
Test-Retest	0 dB	0
Check Fail	0 %	0
ABG > 35	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	∟	∟
NR	↖	↗
Under Masking		

GSI AMTAS™ Hearing Report



QUALIND®		
Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	20
Masker Alert	0 %	0
Trial Time	4.3 s	60
False Alarm	3 %	30
Test-Retest	3 dB	23
Check Fail	0 %	0
ABG > 35	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	1	1

QUALIND®		
Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	20
Masker Alert	0 %	0
Trial Time	4.3 s	60
False Alarm	3 %	30
Test-Retest	3 dB	23
Check Fail	0 %	0
ABG > 35	0	0
ABG < -10	0	0


LEGEND	Right	Left
Masked Air	△	□
Masked Bone	1	1

AMCLASS® Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	MILD-SEVERE	Sloping Hearing Loss	SENSORINEURAL	Symmetric
Left	MILD-SEVERE	Sloping Hearing Loss	SENSORINEURAL	

Speech Recognition Report						
Ear	PTA dB HL	SRT dB HL	Word Recognition (NU6 closed set) Correct % Level dB HL Mask dB HL List			
Right	32	7	92	54	34	NU6-2
Left	23	25	88	45	25	NU6-4

GSI AMTAS™ Hearing Report

AUDIOGRAM Frequency (Hz)



air	15	15	25	30	35	50	Rt
bone		60	75	70	85		Rt
air	10	10	20	25	30	50	Lf
bone		65	65	70	90		Lf

Effective Masking Levels (dB EML)

QUALIND®

Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	30
Masker Alert	0 %	0
Trial Time	3.9 s	65
False Alarm	6 %	50
Test-Retest	0 dB	0
Check Fail	3.2	85
ABG > 35	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	▽	◻

QUALIND®

Quality	GOOD	
Indicators	Value	Rank
Accuracy	6 dB	30
Masker Alert	0 %	0
Trial Time	3.9 s	65
False Alarm	6 %	50
Test-Retest	0 dB	0
Check Fail	3.2	85
ABG > 35	0	0
ABG < -10	0	0

AMCLASS® Audiogram Classification

Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	MILD-SEVERE	Sloping Hearing Loss	SENSORINEURAL	Symmetric
Left	MILD-MODERATE	Sloping Hearing Loss	SENSORINEURAL	


Speech Recognition Report

Ear	PTA dB HL	SRT dB HL	Word Recognition Correct %	Word Recognition (NU6 closed set) Level dB HL	Word Recognition (NU6 closed set) Mask dB HL	Word Recognition (NU6 closed set) List
Right	43	43	88	65	45	NU6-2
Left	38	33	80	60	40	NU6-4

Speech Recognition Report

Ear	PTA dB HL	SRT dB HL	Word Recognition Correct %	Word Recognition (NU6 closed set) Level dB HL	Word Recognition (NU6 closed set) Mask dB HL	Word Recognition (NU6 closed set) List
Right	43	43	88	65	45	NU6-2
Left	38	33	80	60	40	NU6-4

GSI AMTAS™ Hearing Report




QUALIND ®		
Quality	FAIR	
Indicators	Value	Rank
Accuracy	8 dB	50
Masker Alert	0 Z	0
Trial Time	4.6 s	50
False Alarm	10 %	70
Test-Retest	5 dB	46
Check Fail	0 Z	0
ABG > 30	0	0
ABG < -10	1	50

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	▽	◻
NR	u	v

QUALIND ®		
Quality	FAIR	
Indicators	Value	Rank
Accuracy	8 dB	50
Masker Alert	0 Z	0
Trial Time	4.6 s	50
False Alarm	10 %	70
Test-Retest	5 dB	46
Check Fail	0 Z	0
ABG > 30	0	0
ABG < -10	1	50

AMCLASS® Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	NORMAL-MODERATE	Sloping Hearing Loss	SENSORINEURAL	Asymmetric
Left	MILD-SEVERE	Sloping Hearing Loss	SENSORINEURAL	

GSI AMIAS™ Hearing Report



QUALIND®		
Quality	FAIR	
Indicators	Value	Rank
Accuracy	8 dB	60
Masker Alert	0 %	0
Trial Time	5.4 s	95
False Alarm	3 %	25
Test-Retest	5 dB	46
Check Fail	5 %	85
ABG > 35	0	0
ABG < -10	2	80


LEGEND	Right	Left
Masked Air	△	□
Masked Bone	┘	┐

QUALIND®		
Quality	FAIR	
Indicators	Value	Rank
Accuracy	8 dB	60
Masker Alert	0 %	0
Trial Time	5.4 s	95
False Alarm	3 %	25
Test-Retest	5 dB	46
Check Fail	5 %	85
ABG > 35	0	0
ABG < -10	2	80

AMCLASS® Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	NORMAL-MODERATE	Sloping Hearing Loss	SENSORINEURAL	Symmetric
Left	NORMAL-MODERATE	Sloping Hearing Loss	SENSORINEURAL	

GSI AMTAS™ Hearing Report

AUDIOGRAM
Frequency (Hz)



QUALIND®		
Quality	FAIR	
Indicators	Value	Rank
Accuracy	7 dB	50
Masker Alert	9 %	80
Trial Time	3.2 s	15
False Alarm	4 %	35
Test-Retest	0 dB	0
Check Fail	0 %	0
ABG > 35	1	70
ABG < -10	1	50

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	∟	∟
NR	⬇	⬇
Under Masking		


QUALIND®		
Quality	FAIR	
Indicators	Value	Rank
Accuracy	7 dB	50
Masker Alert	9 %	80
Trial Time	3.2 s	15
False Alarm	4 %	35
Test-Retest	0 dB	0
Check Fail	0 %	0
ABG > 35	1	70
ABG < -10	1	50

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	∟	∟
NR	⬇	⬇
Under Masking		

AMCLASS® Audiogram Classification

Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	NORMAL-SEVERE	Sloping Hearing Loss	SENSORINEURAL	Asymmetric
Left	MILD-SEVERE	Sloping Hearing Loss	MIXED	

GSI AMTAS™ Hearing Report



QUALIND®		
Quality	FAIR	
Indicators	Value	Rank
Accuracy	7 dB	45
Masker Alert	0 %	0
Trial Time	3.2 s	15
False Alarm	10 %	65
Test-Retest	18 dB	100
Check Fail	0 %	0
ABG > 35	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	▽	◇

AMCLASS® Audiogram Classification

Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	MILD	Flat Hearing Loss	CONDUCTIVE	Symmetric
Left	NORMAL-MILD	Sloping Hearing Loss	SENSORINEURAL	


Speech Recognition Report

Ear	PTA dB HL	SRT dB HL	Word Recognition Correct %	Word Recognition (NU6 closed set) Level dB HL	Word Recognition (NU6 closed set) Mask dB HL	List NU6-4
Right	20	10	96	42	22	NU6-4
Left	13	8	96	35	15	NU6-3

QUALIND®

Quality	FAIR	
Indicators	Value	Rank
Accuracy	7 dB	45
Masker Alert	0 %	0
Trial Time	3.2 s	15
False Alarm	10 %	65
Test-Retest	18 dB	100
Check Fail	0 %	0
ABG > 35	0	0
ABG < -10	0	0

GSI AMTAS™ Hearing Report



QUALIND®		
Quality	FAIR	
Indicators	Value	Rank
Accuracy	7 dB	45
Masker Alert	0 %	0
Trial Time	3.2 s	15
False Alarm	9 %	65
Test-Retest	5 dB	46
Check Fail	7 %	90
ABG > 35	0	0
ABG < -10	1	50

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	▽	◻


QUALIND®		
Quality	FAIR	
Indicators	Value	Rank
Accuracy	7 dB	45
Masker Alert	0 %	0
Trial Time	3.2 s	15
False Alarm	9 %	65
Test-Retest	5 dB	46
Check Fail	7 %	90
ABG > 35	0	0
ABG < -10	1	50

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	▽	◻

AMCLASS® Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	NORMAL		Cannot determine	Asymmetric
Left	NORMAL		Cannot Determine	

Speech Recognition Report						
Ear	PTA dB HL	SRT dB HL	Word Recognition Correct %	Level dB HL	(NU6 closed set) Mask dB HL	List
Right	3	-2	96	25	5	NU6-3
Left	3	-7	92	25	5	NU6-1

GSI AMTAS™ Hearing Report



QUALIND [®]		
Quality	FAIR	
Indicators	Value	Rank
Accuracy	8 dB	55
Masker Alert	0 %	0
Trial Time	3.7 s	55
False Alarm	14 %	75
Test-Retest	5 dB	46
Check Fail	5 %	85
ABG > 35	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	∟	∟

AMCLASS [®] Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	MILD	Flat Hearing Loss	CONDUCTIVE	Asymmetric
Left	NORMAL-MILD	Sloping Hearing Loss	CONDUCTIVE	

Speech Recognition Report									
Ear	PTA dB HL	SRT dB HL	Word Recognition (NUS closed set)				Word Recognition (NUS closed set)		
			Correct %	Level dB HL	Mask dB HL	List	Correct %	Level dB HL	Mask dB HL
Right	25	20	92	47	27	NUS-4			
Left	12	12	80	34	14	NUS-2	96	42	22


QUALIND [®]		
Quality	FAIR	
Indicators	Value	Rank
Accuracy	8 dB	55
Masker Alert	0 %	0
Trial Time	3.7 s	55
False Alarm	14 %	75
Test-Retest	5 dB	46
Check Fail	5 %	85
ABG > 35	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	∟	∟

Speech Recognition Report									
Ear	PTA dB HL	SRT dB HL	Word Recognition (NUS closed set)				Word Recognition (NUS closed set)		
			Correct %	Level dB HL	Mask dB HL	List	Correct %	Level dB HL	Mask dB HL
Right	25	20	92	47	27	NUS-4			
Left	12	12	80	34	14	NUS-2	96	42	22

GSI AMTAS™ Hearing Report

AUDIOGRAM Frequency (Hz)



air	35	50	65	45	55	65	rt
bone		60	80	80	90		rt
air	55	50	65	50	55	55	lf
bone		80	80	80	95		lf

Effective Masking Levels (dB EML)

QUALIND®		
Quality	P00R	
Indicators	Value	Rank
Accuracy	14 dB	50
Masker Alert	5 %	70
Trial Time	4.2 s	75
False Alarm	0 %	0
Test-Retest	15 dB	100
Check Fail	68 %	100
ABG > 35	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	∟	∟
NR	↖	↗
Under Masking		


AMCLASS® Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	SEVERE	Flat Hearing Loss	MIXED	Asymmetric
Left	SEVERE	Flat Hearing Loss	MIXED	

PTA	
Ear	dB HL
Right	68
Left	70

QUALIND®		
Quality	P00R	
Indicators	Value	Rank
Accuracy	14 dB	50
Masker Alert	5 %	70
Trial Time	4.2 s	75
False Alarm	0 %	0
Test-Retest	15 dB	100
Check Fail	68 %	100
ABG > 35	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	∟	∟
NR	↖	↗
Under Masking		

GSI AMTAS™ Hearing Report



QUALIND®		
Quality	POOR	
Indicators	Value	Rank
Accuracy	18 dB	95
Masker Alert	32 %	90
Trial Time	4.2 s	75
False Alarm	42 %	100
Test-Retest	3 dB	23
Check Fail	0 %	0
ABG > 35	5	93
ABG < -10	6	6


LEGEND		Right	Left
Masked Air		△	□
Masked Bone		▽	◇
Over Masking			
Under Masking			

QUALIND®		
Quality	POOR	
Indicators	Value	Rank
Accuracy	18 dB	95
Masker Alert	32 %	90
Trial Time	4.2 s	75
False Alarm	42 %	100
Test-Retest	3 dB	23
Check Fail	0 %	0
ABG > 35	5	93
ABG < -10	6	6

AMCLASS® Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	MILD-MODERATE	Sloping Hearing Loss	CONDUCTIVE	Asymmetric
Left	MILD-MODERATE	Sloping Hearing Loss	CONDUCTIVE	Asymmetric

GSI AMTAS™ Hearing Report

AUDIOGRAM Frequency (Hz)



air	-5	0	20	35	60	70	80	Rt
bone		55	60	80	95			Rt
air	-10	5	20	30	40	55	80	Lf
bone		50	60	75	15			Lf

Effective Masking Levels (dB EHL)

AMCLASS® Audiogram Classification

Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	MILD-SEVERE	Sloping Hearing Loss	CONDUCTIVE	Symmetric
Left	MILD-SEVERE	Sloping Hearing Loss	MIXED	

Speech Recognition Report

Ear	PTA dB HL	SRT dB HL	Word Recognition (NU6 closed set)				Word Recognition (NU6 closed set)			
			Correct %	Level dB HL	Mask dB HL	List	Correct %	Level dB HL	Mask dB HL	List
Right	30	20	68	52	32	NU6-1	40	60	40	NU6-2
Left	33	23	44	55	35	NU6-4	64	63	43	NU6-3

QUALIND®

Quality	Value	Rank
Accuracy	13 dB	90
Masker Alert	38 %	95
Trial Time	3.8 s	60
False Alarm	19 %	85
Test-Retest	0 dB	0
Check Fail	0 %	0
ABG > 30	0	0
ABG < -10	0	0


LEGEND	Right	Left
Masked Air	△	□
Masked Bone	┘	┐
NR	⋈	⋈
Over Masking		
Under Masking		

QUALIND®

Quality	POOR	
Indicators	Value	Rank
Accuracy	13 dB	90
Masker Alert	38 %	95
Trial Time	3.8 s	60
False Alarm	19 %	85
Test-Retest	0 dB	0
Check Fail	0 %	0
ABG > 30	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	┘	┐
NR	⋈	⋈
Over Masking		
Under Masking		

GSI AMTAS™ Hearing Report



air	-10	-10	-5	-10	-10	-5	rt
bone			40				rt
air	-10	-10	-10	-10	-5	5	lf
bone					45		lf

Effective Masking Levels (dB EML)

AMCLASS® Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	NORMAL		Cannot Determine	Symmetric
Left	NORMAL		Cannot Determine	

Speech Recognition Report						
Ear	PTA dB HL	SRT dB HL	Word Recognition (NU6 closed set) Correct %	Level dB HL	Mask dB HL	List
Right	10	5	92	32	12	NU6-4
Left	8	-2	88	30	10	NU6-1


QUALIND®		
Quality	P00R	
Indicators	Value	Rank
Accuracy	11 dB	85
Masker Alert	0 %	0
Trial Time	4 s	70
False Alarm	14 %	80
Test-Retest	0 dB	0
Check Fail	44 %	100
ABG > 35	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	1	1

QUALIND®		
Quality	P00R	
Indicators	Value	Rank
Accuracy	11 dB	85
Masker Alert	0 %	0
Trial Time	4 s	70
False Alarm	14 %	80
Test-Retest	0 dB	0
Check Fail	44 %	100
ABG > 35	0	0
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	1	1

AMTAS™ Hearing Report



QUALIND®		
Quality	POOR	
Indicators	Value	Rank
Accuracy	12 dB	85
Masker Alert	12 %	85
Trial Time	4 s	70
False Alarm	26 %	90
Test-Retest	3 dB	23
Check Fail	0 %	0
ABG > 35	2	85
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	7	7
Over Masking		

QUALIND®		
Quality	POOR	
Indicators	Value	Rank
Accuracy	12 dB	85
Masker Alert	12 %	85
Trial Time	4 s	70
False Alarm	26 %	90
Test-Retest	3 dB	23
Check Fail	0 %	0
ABG > 35	2	85
ABG < -10	0	0

LEGEND	Right	Left
Masked Air	△	□
Masked Bone	7	7
Over Masking		

AMCLASS® Audiogram Classification				
Ear	Severity	Configuration	Site of Lesion	Bilateral Symmetry
Right	NORMAL-MODERATE	Sloping Hearing Loss	CONDUCTIVE	Asymmetric
Left	NORMAL-MODERATE	Sloping Hearing Loss	CONDUCTIVE	

Speech Recognition Report										
Ear	PTA dB HL	SRT dB HL	Word Recognition (NU6 closed set)				Word Recognition (NU6 closed set)			
			Correct %	Level dB HL	Mask dB HL	List	Correct %	Level dB HL	Mask dB HL	List
Right	3	-2	76	25	5	NU6-2	92	33	13	NU6-3
Left	10	5	84	32	12	NU6-4	84	40	20	NU6-1

AMTAS Automated Audiometry

- Does not replace audiologists.
- Interpretation and intervention requires skillset of experienced audiologist
- Quality indicators give confidence in results



Enhancements in the queue:

- Non-ear specific bone conduction
 - Performed first
 - No masking
 - Reduce errors in responding to masking noise
 - If asymmetry, will re-test bone with masking
- Non-linguistic speech testing
 - Does not require literacy
 - Does not depend on English as first language



DESIGNED
SMART
BUILT
STRONG